

# Required Parts For All PortaCount Calibration

**Sampling Tube and Storage Cap  
must belong to PortaCount  
being tested**



Sampling Tube



**IMPORTANT NOTE:** Storage Cap must be installed for shipping or the PortaCount can be damaged

Storage Cap

*The PortaCount being shipped must have the Storage Cap installed in the unit. It is highly recommended to send PortaCounts to CTC in the manufacturer's case to provide additional protection. Equipment should be packaged properly according to the: Procedure for Packing Equipment to CTC.*